

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031020\
 Data File : VV014807.D
 Acq On : 10 Mar 2020 17:26
 Operator : SY/MD
 Sample : VV0310WBL02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK80

Quant Time: Mar 26 03:42:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM030620WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 11 03:49:31 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	403724	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	391500	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	188092	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	126189	46.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.38%
7) Chloroethane-d5	1.58	69	94369	43.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.38%
11) 1,1-Dichloroethene-d2	2.12	63	151084	33.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.56%
21) 2-Butanone-d5	3.93	46	149251	73.25	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	73.25%
24) Chloroform-d	4.38	84	236091	45.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.28%
26) 1,2-Dichloroethane-d4	5.06	65	162577	46.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.54%
32) Benzene-d6	5.08	84	499249	48.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	6.10	67	154895	45.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.42%
41) Toluene-d8	7.34	98	472578	46.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.36%
43) trans-1,3-Dichloropropene-	7.64	79	71100	42.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.00%
47) 2-Hexanone-d5	8.12	63	149895	98.54	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.54%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	209115	44.37	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.74%
64) 1,2-Dichlorobenzene-d4	11.65	152	188608	50.59	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.18%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031020\
Data File : VV014807.D
Acq On : 10 Mar 2020 17:26
Operator : SY/MD
Sample : VV0310WBL02
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK80

Quant Time: Mar 26 03:42:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM030620WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 11 03:49:31 2020
Response via : Initial Calibration

