

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031023\
 Data File : VV030151.D
 Acq On : 10 Mar 2023 13:39
 Operator : SY/MD
 Sample : 01868-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BD7

Quant Time: Mar 11 00:04:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 11 00:01:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	148756	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	149243	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	74213	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	32300	3.356	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.200%
7) Chloroethane-d5	1.539	69	38321	4.241	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.800%
11) 1,1-Dichloroethene-d2	2.066	63	54727	2.766	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.400%#
20) 2-Butanone-d5	3.828	46	96788	54.540	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.080%
24) Chloroform-d	4.259	84	98736	5.022	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.400%
26) 1,2-Dichloroethane-d4	4.953	65	42841	5.337	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.800%
32) Benzene-d6	4.969	84	162313	4.042	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.800%
36) 1,2-Dichloropropane-d6	5.998	67	57814	4.785	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.800%
41) Toluene-d8	7.252	98	131389	3.452	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.000%#
43) trans-1,3-Dichloroprop...	7.561	79	17071	4.130	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.600%
46) 2-Hexanone-d5	8.037	63	84857	49.624	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.240%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	46850	5.451	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.000%
66) 1,2-Dichlorobenzene-d4	11.570	152	62899	5.009	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.200%
Target Compounds						Qvalue
13) Acetone	2.159	43	2397	1.827	ug/L	99
16) Methylene chloride	2.455	84	6375	0.523	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031023\
Data File : VV030151.D
Acq On : 10 Mar 2023 13:39
Operator : SY/MD
Sample : 01868-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BD7

Quant Time: Mar 11 00:04:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 11 00:01:55 2023
Response via : Initial Calibration

