

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031023\
 Data File : VV030166.D
 Acq On : 10 Mar 2023 19:41
 Operator : SY/MD
 Sample : 01868-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BD3

Quant Time: Mar 11 00:11:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 11 00:01:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	138932	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	141127	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	70935	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	29249	3.254	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	65.000%	
7) Chloroethane-d5	1.539	69	36011	4.267	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	85.400%	
11) 1,1-Dichloroethene-d2	2.066	63	49345	2.670	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	53.400%#	
20) 2-Butanone-d5	3.818	46	96245	58.069	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	116.140%	
24) Chloroform-d	4.262	84	93610	5.098	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.000%	
26) 1,2-Dichloroethane-d4	4.957	65	41669	5.558	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	111.200%	
32) Benzene-d6	4.973	84	155276	4.089	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.800%	
36) 1,2-Dichloropropane-d6	5.998	67	54927	4.808	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.200%	
41) Toluene-d8	7.252	98	120944	3.360	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	67.200%#	
43) trans-1,3-Dichloroprop...	7.564	79	16852	4.312	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.200%	
46) 2-Hexanone-d5	8.034	63	91250	56.431	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	112.860%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	46893	5.770	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	115.400%	
66) 1,2-Dichlorobenzene-d4	11.570	152	59763	4.980	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.600%	
Target Compounds						
					Qvalue	
16) Methylene chloride	2.455	84	3346	0.294	ug/L	89
47) Tetrachloroethene	7.915	164	160514	17.315	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031023\
Data File : VV030166.D
Acq On : 10 Mar 2023 19:41
Operator : SY/MD
Sample : 01868-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BD3

Quant Time: Mar 11 00:11:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 11 00:01:55 2023
Response via : Initial Calibration

