

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031023\
 Data File : VW030158.D
 Acq On : 10 Mar 2023 16:27
 Operator : SY/MD
 Sample : 01868-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BC3

Quant Time: Mar 11 00:07:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 11 00:01:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	165519	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	167051	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	83718	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	34496	3.221	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	64.400%	
7) Chloroethane-d5	1.539	69	43726	4.349	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	87.000%	
11) 1,1-Dichloroethene-d2	2.066	63	60742	2.759	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	55.200%#	
20) 2-Butanone-d5	3.818	46	115494	58.490	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	116.980%	
24) Chloroform-d	4.262	84	112491	5.142	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.800%	
26) 1,2-Dichloroethane-d4	4.957	65	49457	5.538	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	110.800%	
32) Benzene-d6	4.973	84	182552	4.061	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.200%	
36) 1,2-Dichloropropane-d6	5.998	67	65362	4.833	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.600%	
41) Toluene-d8	7.252	98	147633	3.465	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	69.200%#	
43) trans-1,3-Dichloroprop...	7.561	79	19258	4.163	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.200%	
46) 2-Hexanone-d5	8.037	63	107207	56.010	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	112.020%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	56063	5.827	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	116.600%	
66) 1,2-Dichlorobenzene-d4	11.570	152	71560	5.052	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	101.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031023\
Data File : VV030158.D
Acq On : 10 Mar 2023 16:27
Operator : SY/MD
Sample : 01868-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BC3

Quant Time: Mar 11 00:07:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 11 00:01:55 2023
Response via : Initial Calibration

