

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031124\
 Data File : VW034535.D
 Acq On : 11 Mar 2024 11:55
 Operator : SY/MD
 Sample : P1637-08
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCNR1

Quant Time: Mar 12 01:31:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 12 01:30:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	320770	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	300369	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	142466	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	97716	35.140	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	70.280%
7) Chloroethane-d5	1.529	69	88343	40.779	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.560%
11) 1,1-Dichloroethene-d2	2.056	65	46312	36.826	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.660%
21) 2-Butanone-d5	3.777	46	183142	98.485	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.490%
24) Chloroform-d	4.246	84	214175	43.561	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.120%
26) 1,2-Dichloroethane-d4	4.937	65	149196	46.436	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.880%
32) Benzene-d6	4.957	84	395103	43.585	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.160%
36) 1,2-Dichloropropane-d6	5.985	67	128352	46.395	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.780%
41) Toluene-d8	7.239	98	335436	40.889	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.780%
43) trans-1,3-Dichloroprop...	7.551	79	57893	41.672	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.340%
47) 2-Hexanone-d5	8.021	63	133131	96.368	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.370%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	177439	48.737	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.480%
66) 1,2-Dichlorobenzene-d4	11.558	152	134090	46.047	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.100%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031124\
Data File : VV034535.D
Acq On : 11 Mar 2024 11:55
Operator : SY/MD
Sample : P1637-08
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCNR1

Quant Time: Mar 12 01:31:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 12 01:30:00 2024
Response via : Initial Calibration

