

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031124\
 Data File : VV034546.D
 Acq On : 11 Mar 2024 16:29
 Operator : SY/MD
 Sample : P1637-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Mar 12 01:34:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 12 01:30:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	314975	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	292382	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	139564	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	93226	34.142	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.280%
7) Chloroethane-d5	1.526	69	86330	40.583	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.160%
11) 1,1-Dichloroethene-d2	2.056	65	44033	35.658	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.320%
21) 2-Butanone-d5	3.780	46	204088	111.768	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.770%
24) Chloroform-d	4.246	84	206473	42.767	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.540%
26) 1,2-Dichloroethane-d4	4.941	65	145146	46.007	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.020%
32) Benzene-d6	4.957	84	376806	42.702	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.400%
36) 1,2-Dichloropropane-d6	5.986	67	126328	46.910	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.820%
41) Toluene-d8	7.243	98	321148	40.217	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.440%
43) trans-1,3-Dichloroprop...	7.551	79	57078	42.207	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.420%
47) 2-Hexanone-d5	8.021	63	140642	104.586	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.590%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	177762	50.160	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.320%
66) 1,2-Dichlorobenzene-d4	11.561	152	131296	46.025	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.040%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031124\
Data File : VV034546.D
Acq On : 11 Mar 2024 16:29
Operator : SY/MD
Sample : P1637-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Mar 12 01:34:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 12 01:30:00 2024
Response via : Initial Calibration

