

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031220\
 Data File : VV014883.D
 Acq On : 12 Mar 2020 21:28
 Operator : SY/MD
 Sample : L1890-04
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0D87

Quant Time: Mar 13 05:13:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM030620WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 13 04:10:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	400379	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	385654	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	184754	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	101170	37.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.68%
7) Chloroethane-d5	1.58	69	82262	38.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	76.80%
11) 1,1-Dichloroethene-d2	2.12	63	133603	29.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.34%#
21) 2-Butanone-d5	3.93	46	154257	76.33	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	76.33%
24) Chloroform-d	4.38	84	231761	45.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.34%
26) 1,2-Dichloroethane-d4	5.06	65	152496	43.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.52%
32) Benzene-d6	5.08	84	465726	45.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.48%
36) 1,2-Dichloropropane-d6	6.10	67	142728	42.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.58%
41) Toluene-d8	7.34	98	442006	44.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.64%
43) trans-1,3-Dichloropropene-	7.64	79	59776	36.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.56%
47) 2-Hexanone-d5	8.12	63	139504	93.09	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.09%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	202715	43.67	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.34%
64) 1,2-Dichlorobenzene-d4	11.65	152	179960	49.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.28%

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031220\
 Data File : VV014883.D
 Acq On : 12 Mar 2020 21:28
 Operator : SY/MD
 Sample : L1890-04
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampled :
 C0D87

Quant Time: Mar 13 05:13:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM030620WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 13 04:10:22 2020
 Response via : Initial Calibration

