

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031224\
 Data File : VW034552.D
 Acq On : 12 Mar 2024 10:24
 Operator : SY/MD
 Sample : P1715-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
ClientSampleId :

Quant Time: Mar 13 04:57:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSI030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 04:17:25 2024
 Response via : Initial Calibration

Compound	R. T.	QI on	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	10631	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9845	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	5102	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	5985	0.525	ug/L	0.00
Spike Amount	0.500	Range	40 - 130	Recovery	=	106.000%
4) 1,2-Dichloroethane-d4	4.969	65	3644	0.516	ug/L	0.00
Spike Amount	0.500	Range	70 - 130	Recovery	=	104.000%
7) 1,2-Dichloropropane-d6	5.995	67	3924	0.513	ug/L	0.00
Spike Amount	0.500	Range	60 - 140	Recovery	=	102.000%
8) Toluene-d8	7.253	98	11220	0.495	ug/L	0.00
Spike Amount	0.500	Range	70 - 130	Recovery	=	98.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2487	0.453	ug/L	0.00
Spike Amount	0.500	Range	65 - 120	Recovery	=	90.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031224\
 Data File : VV034552.D
 Acq On : 12 Mar 2024 10:24
 Operator : SY/MD
 Sample : P1715-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
ClientSampleId :

Quant Time: Mar 13 04:57:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSI M030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 04:17:25 2024
 Response via : Initial Calibration

