

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031224\
 Data File : VV034555.D
 Acq On : 12 Mar 2024 11:35
 Operator : SY/MD
 Sample : P1715-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Mar 13 04:57:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 04:17:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	11011	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	10190	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	5307	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	6360	0.539	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	108.000%
4) 1,2-Dichloroethane-d4	4.969	65	3987	0.545	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	110.000%
7) 1,2-Dichloropropane-d6	5.995	67	4225	0.533	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	106.000%
8) Toluene-d8	7.253	98	11942	0.509	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2877	0.507	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%
Target Compounds						
12) 1,2,3-Trichloropropane	10.317	75	100	0.022	ug/L	# 77


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031224\
Data File : VV034555.D
Acq On : 12 Mar 2024 11:35
Operator : SY/MD
Sample : P1715-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :


Quant Time: Mar 13 04:57:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 04:17:25 2024
Response via : Initial Calibration





#12
 1,2,3-Trichloropropane
 Concen: 0.022 ug/L
 RT: 10.317 min Scan# 61
 Delta R.T. 0.103 min
 Lab File: VV034555.D
 Acq: 12 Mar 2024 11:35

Instrument :
 MSVOA_V
 ClientSampleId :



Tgt Ion: 75 Resp: 100

Ion	Ratio	Lower	Upper
75	100		
110	37.0	33.1	49.7
77	8.0	26.2	39.4

