

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031224\
 Data File : VV034556.D
 Acq On : 12 Mar 2024 11:59
 Operator : SY/MD
 Sample : P1715-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Mar 13 04:57:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 04:17:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	11560	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.801	117	10585	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.192	152	5547	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.281	65	6292	0.508	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	102.000%
4) 1,2-Dichloroethane-d4	4.970	65	4049	0.527	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.000%
7) 1,2-Dichloropropane-d6	5.996	67	4222	0.513	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	102.000%
8) Toluene-d8	7.254	98	11816	0.485	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2876	0.488	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	98.000%
Target Compounds						
12) 1,2,3-Trichloropropane	10.317	75	60	0.013	ug/L	Qvalue # 85


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031224\
Data File : VV034556.D
Acq On : 12 Mar 2024 11:59
Operator : SY/MD
Sample : P1715-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :


Quant Time: Mar 13 04:57:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 04:17:25 2024
Response via : Initial Calibration





#12
 1,2,3-Trichloropropane
 Concen: 0.013 ug/L
 RT: 10.317 min Scan# 61
 Delta R.T. 0.103 min
 Lab File: VV034556.D
 Acq: 12 Mar 2024 11:59

Instrument :
 MSVOA_V
 ClientSampleId :



Tgt Ion: 75 Resp: 60

Ion	Ratio	Lower	Upper
75	100		
110	40.0	33.1	49.7
77	15.0	26.2	39.4

