

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031224\
 Data File : VV034549.D
 Acq On : 12 Mar 2024 09:00
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5284

Quant Time: Mar 13 04:54:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 04:17:25 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	11154	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.800	117	10414	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	5418	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	6888	0.576	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	116.000%
4) 1,2-Dichloroethane-d4	4.945	65	4421	0.597	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	120.000%
7) 1,2-Dichloropropane-d6	5.995	67	4661	0.576	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	116.000%
8) Toluene-d8	7.253	98	13165	0.549	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	110.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3348	0.577	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	116.000%
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	6748	0.542	ug/L	100
6) Trichloroethene	5.852	95	4325	0.530	ug/L	99
9) 1,2-Dibromoethane	8.290	107	2733	0.553	ug/L	98
12) 1,2,3-Trichloropropane	10.197	75	2511	0.552	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.356	75	510	0.524	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031224\
Data File : VV034549.D
Acq On : 12 Mar 2024 09:00
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5284

Quant Time: Mar 13 04:54:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 04:17:25 2024
Response via : Initial Calibration

