

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031323\
 Data File : VW030179.D
 Acq On : 13 Mar 2023 16:15
 Operator : SY/MD
 Sample : 01884-03DL 200X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CB908DL

Quant Time: Mar 14 03:46:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Mar 14 03:44:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	133783	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.796	117	137088	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	69510	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	54156	6.256	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	125.200%	
7) Chloroethane-d5	1.539	69	46440	5.715	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	114.400%	
11) 1,1-Dichloroethene-d2	2.066	63	77852	4.375	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	87.400%	
20) 2-Butanone-d5	3.828	46	76365	47.848	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	95.700%	
24) Chloroform-d	4.265	84	94665	5.353	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	107.000%	
26) 1,2-Dichloroethane-d4	4.956	65	40304	5.583	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	111.600%	
32) Benzene-d6	4.973	84	176746	4.791	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.800%	
36) 1,2-Dichloropropane-d6	6.001	67	55134	4.968	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.400%	
41) Toluene-d8	7.252	98	153903	4.402	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.000%	
43) trans-1,3-Dichloroprop...	7.564	79	18751	4.939	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	98.800%	
46) 2-Hexanone-d5	8.037	63	66751	42.496	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	85.000%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	40141	5.084	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.600%	
66) 1,2-Dichlorobenzene-d4	11.570	152	61234	5.207	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	104.200%	
Target Compounds						
						Qvalue
16) Methylene chloride	2.455	84	2700	0.246	ug/L	91
51) Chlorobenzene	8.824	112	5760	0.211	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031323\
 Data File : VV030179.D
 Acq On : 13 Mar 2023 16:15
 Operator : SY/MD
 Sample : 01884-03DL 200X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CB908DL

Quant Time: Mar 14 03:46:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Mar 14 03:44:35 2023
 Response via : Initial Calibration

