

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031323\
 Data File : VW030173.D
 Acq On : 13 Mar 2023 13:53
 Operator : SY/MD
 Sample : 01838-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK002

Quant Time: Mar 14 03:45:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Mar 14 03:44:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	146591	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	146417	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	72942	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	56984	6.008	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.200%
7) Chloroethane-d5	1.536	69	48202	5.414	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.200%
11) 1,1-Dichloroethene-d2	2.063	63	83734	4.294	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.800%
20) 2-Butanone-d5	3.847	46	72379	41.388	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.780%
24) Chloroform-d	4.262	84	99389	5.130	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.600%
26) 1,2-Dichloroethane-d4	4.957	65	42622	5.389	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.800%
32) Benzene-d6	4.970	84	191443	4.859	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
36) 1,2-Dichloropropane-d6	5.998	67	58815	4.962	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.200%
41) Toluene-d8	7.252	98	166698	4.464	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.200%
43) trans-1,3-Dichloroprop...	7.561	79	20593	5.078	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.600%
46) 2-Hexanone-d5	8.037	63	70574	42.068	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.140%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	42135	4.997	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.000%
66) 1,2-Dichlorobenzene-d4	11.571	152	62860	5.093	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031323\
Data File : VV030173.D
Acq On : 13 Mar 2023 13:53
Operator : SY/MD
Sample : 01838-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK002

Quant Time: Mar 14 03:45:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Mar 14 03:44:35 2023
Response via : Initial Calibration

