



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV031325\
Data File : VV038696.D
Acq On : 13 Mar 2025 09:34
Operator : SY/MD
Sample : BFB206
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031325WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Thu Mar 13 13:15:04 2025



AutoFind: Scans 2787, 2788, 2789; Background Corrected with Scan 2776

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.6	14207	PASS
75	95	30	80	50.7	40893	PASS
95	95	100	100	100.0	80579	PASS
96	95	5	9	6.4	5186	PASS
173	174	0.00	2	1.8	1086	PASS
174	95	50	120	76.8	61901	PASS
175	174	5	9	7.5	4621	PASS
176	174	95	105	97.0	60037	PASS
177	176	5	10	5.9	3523	PASS

m/z	Abundance
40.00	234.0
44.00	389.0

BEESON
Client Sample :
MS/MS
Instrument :

m/z	Abundance
36.00	664.0
37.00	3649.0
38.10	3354.0
39.00	1541.0
40.10	323.0
40.70	111.0
44.10	928.0
44.90	763.0
47.00	997.0
47.90	551.0
49.00	2984.0
50.00	14162.0
51.00	4556.0
52.00	331.0
53.00	133.0
53.90	147.0
55.00	243.0
56.00	1174.0
57.00	2340.0
60.00	879.0
61.00	3940.0
62.00	4205.0
63.00	3782.0
64.10	772.0
65.00	559.0
66.00	110.0
66.80	267.0
68.00	8199.0
69.00	8068.0
70.00	610.0
71.90	507.0
73.10	3575.0
74.10	13779.0
75.00	40032.0
76.10	3535.0
77.00	418.0
77.90	285.0
79.00	4162.0
79.90	1087.0
80.90	3898.0
81.90	952.0
87.00	3137.0
88.00	4116.0
88.80	107.0
90.90	329.0
92.00	2612.0
93.00	3653.0
94.00	9712.0
95.00	79408.0
96.10	5368.0
97.00	176.0
104.00	421.0
104.80	117.0
105.90	421.0
110.80	106.0
112.80	117.0
115.20	132.0
115.90	412.0
116.90	798.0
117.90	454.0
118.90	570.0
127.90	281.0
129.00	180.0
129.90	363.0
130.90	157.0
134.80	210.0
136.90	177.0
140.90	900.0
143.00	1444.0
144.90	624.0
146.00	115.0
148.00	255.0
151.90	101.0
154.90	275.0
170.90	328.0
171.90	819.0
173.10	1046.0
174.00	57792.0
175.00	4550.0
176.00	56448.0
177.00	3466.0
178.00	140.0

BEBS04
ClientSampleId :
MS2A047A
Instrument :

m/z	Abundance
36.00	776.0
37.00	3772.0
38.00	3494.0
39.00	1703.0
40.10	349.0
40.70	139.0
44.10	831.0
44.90	719.0
47.00	1136.0
47.90	495.0
49.00	3010.0
50.00	14663.0
51.00	4495.0
52.00	316.0
53.00	156.0
53.80	153.0
55.00	314.0
56.00	1246.0
57.10	2325.0
58.10	111.0
60.00	907.0
61.00	4120.0
62.00	4357.0
63.00	4078.0
64.00	674.0
65.10	659.0
66.10	127.0
66.90	314.0
68.00	8448.0
69.00	8658.0
70.00	671.0
71.90	480.0
73.10	3705.0
74.00	14299.0
75.00	42432.0
76.10	3398.0
77.00	490.0
77.90	324.0
79.00	4317.0
79.90	1261.0
80.90	4069.0
82.00	1061.0
85.90	166.0
87.00	3272.0
88.00	4319.0
88.80	129.0
90.90	389.0
92.00	2742.0
93.00	3616.0
94.00	10148.0
95.10	83216.0
96.10	5345.0
97.00	192.0
104.00	417.0
104.80	145.0
105.90	394.0
108.00	117.0
110.80	115.0
112.90	100.0
115.10	123.0
116.00	366.0
116.90	817.0
117.90	450.0
118.90	622.0
127.90	293.0
129.00	185.0
130.00	381.0
130.90	162.0
134.80	258.0
137.00	166.0
141.00	993.0
143.00	1385.0
144.90	606.0
145.90	164.0
147.90	241.0
151.90	140.0
154.90	342.0
171.00	412.0
171.90	998.0
173.00	1136.0
174.00	64024.0
175.00	4817.0
176.00	62464.0
177.00	3565.0
177.90	137.0

BEBS04
ClientSampleId :
MS/MS
Instrument :

m/z	Abundance
36.00	774.0
37.00	3468.0
38.00	3157.0
39.00	1606.0
40.00	307.0
40.70	132.0
44.00	718.0
45.00	668.0
47.00	1098.0
47.90	379.0
49.00	2870.0
50.00	13796.0
51.00	4080.0
52.00	291.0
53.00	165.0
53.70	120.0
55.10	332.0
56.00	1151.0
57.00	2135.0
60.00	886.0
61.00	3869.0
62.00	4035.0
63.00	3893.0
64.00	628.0
65.10	757.0
66.10	127.0
66.90	329.0
68.00	7952.0
69.00	8191.0
70.10	613.0
71.90	405.0
73.00	3565.0
74.10	13364.0
75.00	40216.0
76.10	2984.0
77.00	524.0
77.90	386.0
79.00	4094.0
79.90	1277.0
81.00	3945.0
82.00	1060.0
85.90	221.0
87.00	3258.0
88.00	4090.0
90.90	400.0
92.00	2661.0
93.00	3331.0
94.00	9837.0
95.00	79112.0
96.10	4846.0
97.00	200.0
104.00	338.0
104.80	159.0
105.90	334.0
108.10	134.0
110.80	130.0
113.00	107.0
116.00	333.0
116.90	744.0
117.90	437.0
118.90	580.0
128.00	315.0
129.00	160.0
130.00	291.0
130.90	112.0
134.90	248.0
137.00	169.0
141.00	957.0
143.00	1282.0
144.90	568.0
145.90	167.0
147.80	217.0
151.90	126.0
154.90	329.0
171.00	394.0
171.90	1065.0
173.00	1075.0
174.00	63888.0
175.00	4495.0
176.00	61200.0
177.00	3537.0
177.90	120.0

BEBS04
ClientSampleId :
MS2A047A
Instrument :