

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031423\
 Data File : VV030188.D
 Acq On : 14 Mar 2023 11:46
 Operator : SY/MD
 Sample : VV0314WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK311

Quant Time: Mar 15 01:37:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 15 01:37:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	134508	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.796	117	138013	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	66392	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	52240	6.002	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 120.000%		
7) Chloroethane-d5	1.536	69	44913	5.497	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 110.000%		
11) 1,1-Dichloroethene-d2	2.063	63	76442	4.272	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 85.400%		
20) 2-Butanone-d5	3.851	46	69469	43.293	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	= 86.580%		
24) Chloroform-d	4.262	84	93882	5.281	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 105.600%		
26) 1,2-Dichloroethane-d4	4.957	65	40821	5.624	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 112.400%		
32) Benzene-d6	4.973	84	173132	4.662	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 93.200%		
36) 1,2-Dichloropropane-d6	6.002	67	55725	4.988	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 99.800%		
41) Toluene-d8	7.252	98	148468	4.218	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 84.400%		
43) trans-1,3-Dichloroprop...	7.564	79	19198	5.023	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 100.400%		
46) 2-Hexanone-d5	8.040	63	69070	43.678	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 87.360%		
56) 1,1,2,2-Tetrachloroeth...	10.162	84	40565	5.104	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 102.000%		
66) 1,2-Dichlorobenzene-d4	11.571	152	56543	5.034	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 100.600%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031423\
Data File : VV030188.D
Acq On : 14 Mar 2023 11:46
Operator : SY/MD
Sample : VV0314WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK311

Quant Time: Mar 15 01:37:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Mar 15 01:37:04 2023
Response via : Initial Calibration

