

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031521\
 Data File : VW020600.D
 Acq On : 15 Mar 2021 12:19
 Operator : SY/MD
 Sample : M1655-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BB2

Quant Time: Mar 16 03:44:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM031321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 16 03:42:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	429833	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	421825	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	178767	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	137191	39.57	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	79.14%		
7) Chloroethane-d5	1.568	69	127574	45.99	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	91.98%		
11) 1,1-Dichloroethene-d2	2.108	63	211592	32.10	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	64.20%		
21) 2-Butanone-d5	3.886	46	239029	107.42	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	107.42%		
24) Chloroform-d	4.352	84	286071	47.47	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.94%		
26) 1,2-Dichloroethane-d4	5.037	65	200907	50.53	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	101.06%		
32) Benzene-d6	5.053	84	555314	47.65	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	95.30%		
36) 1,2-Dichloropropane-d6	6.072	67	188186	50.47	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	100.94%		
41) Toluene-d8	7.320	98	460133	44.42	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	88.84%		
43) trans-1,3-Dichloroprop...	7.625	79	63226	36.28	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	72.56%		
47) 2-Hexanone-d5	8.091	63	130618	93.63	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	93.63%		
57) 1,1,2,2-Tetrachloroeth...	10.220	84	228493	49.04	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	98.08%		
64) 1,2-Dichlorobenzene-d4	11.632	152	178257	51.46	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	102.92%		
Target Compounds						Qvalue
25) Chloroform	4.381	83	12216	1.98	ug/L	97
48) 2-Hexanone	8.091	43	18305	4.69	ug/L #	85
59) 1,2,3-Trichloropropane	11.255	75	25455	6.04	ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031521\
Data File : VV020600.D
Acq On : 15 Mar 2021 12:19
Operator : SY/MD
Sample : M1655-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BB2

Quant Time: Mar 16 03:44:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM031321WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 16 03:42:12 2021
Response via : Initial Calibration









#25
Chloroform
Concen: 1.98 ug/L
RT: 4.381 min Scan# 1039
Delta R.T. 0.009 min
Lab File: VV020600.D
Acq: 15 Mar 2021 12:19




Instrument :
MSVOA_V
ClientSampled :
C0BB2


Tgt Ion: 83 Resp: 12216
Ion Ratio Lower Upper
83 100
85 65.7 44.2 82.2



#48
2-Hexanone
Concen: 4.69 ug/L
RT: 8.091 min Scan# 2193
Delta R.T. -0.048 min
Lab File: VV020600.D
Acq: 15 Mar 2021 12:19

Tgt Ion: 43 Resp: 18305
Ion Ratio Lower Upper
43 100
58 44.6 41.8 62.6
57 27.5 15.0 22.4#
100 1.7 8.9 13.3#





#59
 1,2,3-Trichloropropane
 Concen: 6.04 ug/L
 RT: 11.255 min Scan# 3177
 Delta R.T. 0.977 min
 Lab File: VV020600.D
 Acq: 15 Mar 2021 12:19

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BB2

Tgt Ion: 75 Resp: 25455
 Ion Ratio Lower Upper
 75 100
 77 33.9 25.4 38.0

