

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031523\
 Data File : VW030212.D
 Acq On : 15 Mar 2023 15:24
 Operator : SY/MD
 Sample : 01884-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CB919

Manual Integrations
 APPROVED

Reviewed By :Krupa
 Patel

Quant Time: Mar 16 01:52:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 16 01:47:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	150313	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	147480	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	76977	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	48650	5.002	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	100.000%	
7) Chloroethane-d5	1.539	69	44097	4.830	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.600%	
11) 1,1-Dichloroethene-d2	2.066	63	73107	3.656	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	73.200%	
20) 2-Butanone-d5	3.821	46	70266	39.185	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	78.380%	
24) Chloroform-d	4.259	84	95265	4.795	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.800%	
26) 1,2-Dichloroethane-d4	4.953	65	39581	4.880	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.600%	
32) Benzene-d6	4.973	84	179630	4.527	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.600%	
36) 1,2-Dichloropropane-d6	5.998	67	55951	4.687	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.800%	
41) Toluene-d8	7.252	98	153322	4.076	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.600%	
43) trans-1,3-Dichloroprop...	7.561	79	18450	4.517	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.400%	
46) 2-Hexanone-d5	8.037	63	63087	37.334	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	74.660%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	39378	4.636	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	92.800%	
66) 1,2-Dichlorobenzene-d4	11.570	152	63775	4.897	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	98.000%	
Target Compounds						
13) Acetone	2.156	43	6100m	4.603	ug/L	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031523\
Data File : VV030212.D
Acq On : 15 Mar 2023 15:24
Operator : SY/MD
Sample : 01884-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CB919

Manual Integrations
APPROVED

Reviewed By :Krupa
Patel

Quant Time: Mar 16 01:52:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 16 01:47:11 2023
Response via : Initial Calibration

