

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031524\
 Data File : VW034677.D
 Acq On : 16 Mar 2024 01:04
 Operator : SY/MD
 Sample : P1740-16
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0GK1

Quant Time: Mar 16 01:45:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 16 00:58:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	347197	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	328012	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	147520	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	89182	29.630	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	59.260%#	
7) Chloroethane-d5	1.529	69	84270	35.938	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	71.880%	
11) 1,1-Dichloroethene-d2	2.059	65	44848	32.948	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	65.900%	
21) 2-Butanone-d5	3.786	46	186413	92.614	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	92.610%	
24) Chloroform-d	4.249	84	212818	39.991	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	79.980%	
26) 1,2-Dichloroethane-d4	4.940	65	142349	40.933	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	81.860%	
32) Benzene-d6	4.960	84	393110	39.710	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	79.420%	
36) 1,2-Dichloropropane-d6	5.989	67	127341	42.150	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	84.300%	
41) Toluene-d8	7.243	98	340714	38.032	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	76.060%#	
43) trans-1,3-Dichloroprop...	7.554	79	56523	37.257	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	74.520%	
47) 2-Hexanone-d5	8.024	63	131113	86.909	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	86.910%	
56) 1,1,2,2-Tetrachloroeth...	10.152	84	183165	46.070	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	92.140%	
66) 1,2-Dichlorobenzene-d4	11.561	152	133620	44.313	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	88.620%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031524\
Data File : VV034677.D
Acq On : 16 Mar 2024 01:04
Operator : SY/MD
Sample : P1740-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0GK1

Quant Time: Mar 16 01:45:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 16 00:58:04 2024
Response via : Initial Calibration

