

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031524\
 Data File : VV034685.D
 Acq On : 16 Mar 2024 04:16
 Operator : SY/MD
 Sample : P1742-22
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0761

Quant Time: Mar 16 05:21:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 16 00:58:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	689191	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	640606	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	289095	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	93333	15.622	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	31.240%#
7) Chloroethane-d5	1.529	69	88060	18.919	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	37.840%#
11) 1,1-Dichloroethene-d2	2.056	65	44766	16.568	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	33.140%#
21) 2-Butanone-d5	3.789	46	180264	45.118	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	45.120%
24) Chloroform-d	4.249	84	212838	20.148	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	40.300%#
26) 1,2-Dichloroethane-d4	4.940	65	143965	20.855	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	41.720%#
32) Benzene-d6	4.960	84	393654	20.361	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	40.720%#
36) 1,2-Dichloropropane-d6	5.989	67	128386	21.759	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	43.520%#
41) Toluene-d8	7.243	98	339302	19.393	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	38.780%#
43) trans-1,3-Dichloroprop...	7.554	79	52272	17.642	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	35.280%#
47) 2-Hexanone-d5	8.024	63	124251	42.171	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	42.170%#
56) 1,1,2,2-Tetrachloroeth...	10.152	84	171749	22.119	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	44.240%#
66) 1,2-Dichlorobenzene-d4	11.561	152	131543	22.261	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	44.520%#
Target Compounds						
25) Chloroform	4.278	83	39705	3.915	ug/L	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031524\
Data File : VV034685.D
Acq On : 16 Mar 2024 04:16
Operator : SY/MD
Sample : P1742-22
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0761

Quant Time: Mar 16 05:21:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 16 00:58:04 2024
Response via : Initial Calibration

