

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031524\
 Data File : VW034661.D
 Acq On : 15 Mar 2024 18:17
 Operator : SY/MD
 Sample : P1766-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COGM5

Quant Time: Mar 15 23:36:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 15 23:30:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	368889	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	351263	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	157110	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	92091	28.798	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	57.600%#
7) Chloroethane-d5	1.532	69	89071	35.752	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	71.500%
11) 1,1-Dichloroethene-d2	2.059	65	46534	32.176	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.360%
21) 2-Butanone-d5	3.793	46	202121	94.513	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.510%
24) Chloroform-d	4.246	84	221487	39.172	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.340%
26) 1,2-Dichloroethane-d4	4.940	65	149817	40.547	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.100%
32) Benzene-d6	4.960	84	412672	38.927	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.860%
36) 1,2-Dichloropropane-d6	5.989	67	134629	41.613	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.220%
41) Toluene-d8	7.243	98	353367	36.834	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	73.660%#
43) trans-1,3-Dichloroprop...	7.554	79	58214	35.832	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.660%
47) 2-Hexanone-d5	8.024	63	142428	88.160	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.160%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	189787	44.576	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.160%
66) 1,2-Dichlorobenzene-d4	11.561	152	141918	44.192	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031524\
Data File : VV034661.D
Acq On : 15 Mar 2024 18:17
Operator : SY/MD
Sample : P1766-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0GM5

Quant Time: Mar 15 23:36:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 15 23:30:32 2024
Response via : Initial Calibration

