

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031619\
 Data File : VV009689.D
 Acq On : 15 Mar 2019 16:39
 Operator : SY/MD
 Sample : K1939-06
 Misc : 5.59g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BF0S9

Quant Time: Mar 16 02:57:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 16 02:32:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	164718	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	164624	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	71079	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	39855	40.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.76%
7) Chloroethane-d5	1.55	69	24400	24.05	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	48.10%#
11) 1,1-Dichloroethene-d2	2.10	63	86359	33.58	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.16%
21) 2-Butanone-d5	3.95	46	72080	100.05	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.05%
24) Chloroform-d	4.40	84	98101	46.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.14%
26) 1,2-Dichloroethane-d4	5.08	65	67632	48.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.00%
32) Benzene-d6	5.09	84	199207	46.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.08%
36) 1,2-Dichloropropane-d6	6.12	67	62350	48.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.42%
41) Toluene-d8	7.36	98	194578	46.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.32%
43) trans-1,3-Dichloropropene-	7.67	79	28557	43.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.40%
47) 2-Hexanone-d5	8.14	63	60269	107.65	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.65%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	99545	48.41	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.82%
64) 1,2-Dichlorobenzene-d4	11.68	152	72740	48.51	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.02%

Target Compounds

					Ovalue
20) cis-1,2-Dichloroethene	3.96	96	24798	21.910 ug/L	97
34) Trichloroethene	5.96	95	203224	184.288 ug/L	98
46) Tetrachloroethene	8.02	164	16178	18.660 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031619\
Data File : VV009689.D
Acq On : 15 Mar 2019 16:39
Operator : SY/MD
Sample : K1939-06
Misc : 5.59g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF0S9

Quant Time: Mar 16 02:57:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031219WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 16 02:32:06 2019
Response via : Initial Calibration

