

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031619\
 Data File : VV009690.D
 Acq On : 15 Mar 2019 17:03
 Operator : SY/MD
 Sample : K1939-07
 Misc : 5.85g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BF0T0

Quant Time: Mar 16 03:01:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 16 02:32:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	146910	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	147434	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	65944	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	35535	40.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.74%
7) Chloroethane-d5	1.55	69	23620	26.10	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	52.20%#
11) 1,1-Dichloroethene-d2	2.10	63	82761	36.08	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.16%
21) 2-Butanone-d5	3.95	46	67082	104.40	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.40%
24) Chloroform-d	4.40	84	88923	46.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.64%
26) 1,2-Dichloroethane-d4	5.08	65	61695	49.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.18%
32) Benzene-d6	5.09	84	184931	48.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.48%
36) 1,2-Dichloropropane-d6	6.12	67	58027	50.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.24%
41) Toluene-d8	7.36	98	171866	46.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.04%
43) trans-1,3-Dichloropropene-	7.67	79	26901	45.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.86%
47) 2-Hexanone-d5	8.14	63	53109	105.93	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.93%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	90744	49.28	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.56%
64) 1,2-Dichlorobenzene-d4	11.67	152	66643	47.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.80%

Target Compounds					Ovalue
20) cis-1,2-Dichloroethene	3.96	96	28507	28.241 ug/L	96
34) Trichloroethene	5.96	95	31002	31.391 ug/L	98
46) Tetrachloroethene	8.02	164	9644	12.421 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031619\
Data File : VV009690.D
Acq On : 15 Mar 2019 17:03
Operator : SY/MD
Sample : K1939-07
Misc : 5.85g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF0T0

Quant Time: Mar 16 03:01:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031219WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 16 02:32:06 2019
Response via : Initial Calibration

