

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031619\
 Data File : VV009708.D
 Acq On : 16 Mar 2019 00:50
 Operator : SY/MD
 Sample : VV0316MBL02
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK73

Quant Time: Mar 16 05:21:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 16 02:32:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	151915	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	147883	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	61827	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	43618	48.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.02%
7) Chloroethane-d5	1.58	69	43638	46.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.26%
11) 1,1-Dichloroethene-d2	2.13	63	92906	39.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.34%
21) 2-Butanone-d5	3.93	46	67071	100.94	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.94%
24) Chloroform-d	4.41	84	95146	48.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.88%
26) 1,2-Dichloroethane-d4	5.08	65	65498	50.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.80%
32) Benzene-d6	5.10	84	196161	51.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.02%
36) 1,2-Dichloropropane-d6	6.12	67	60145	52.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.62%
41) Toluene-d8	7.36	98	185489	49.51	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.02%
43) trans-1,3-Dichloropropene-	7.67	79	25805	43.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.90%
47) 2-Hexanone-d5	8.14	63	46273	92.01	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.01%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	83616	45.27	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.54%
64) 1,2-Dichlorobenzene-d4	11.68	152	64592	49.52	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.04%

Target Compounds					Ovalue
46) Tetrachloroethene	8.02	164	4021	5.163 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031619\
Data File : VV009708.D
Acq On : 16 Mar 2019 00:50
Operator : SY/MD
Sample : VV0316MBL02
Misc : 5.00µL/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK73

Quant Time: Mar 16 05:21:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031219WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 16 02:32:06 2019
Response via : Initial Calibration

