

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031621\
 Data File : VV020631.D
 Acq On : 16 Mar 2021 15:26
 Operator : SY/MD
 Sample : M1675-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BC9

Quant Time: Mar 17 01:51:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM031321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 17 01:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	365192	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	362110	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	155901	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	108309	36.77	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	73.54%		
7) Chloroethane-d5	1.571	69	101620	43.12	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	86.24%		
11) 1,1-Dichloroethene-d2	2.108	63	168998	30.18	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	60.36%		
21) 2-Butanone-d5	3.886	46	229697	121.49	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	121.49%		
24) Chloroform-d	4.352	84	247597	48.36	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	96.72%		
26) 1,2-Dichloroethane-d4	5.037	65	179339	53.09	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	106.18%		
32) Benzene-d6	5.053	84	487455	48.72	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.44%		
36) 1,2-Dichloropropane-d6	6.072	67	169249	52.88	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	105.76%		
41) Toluene-d8	7.320	98	406931	45.76	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	91.52%		
43) trans-1,3-Dichloroprop...	7.625	79	73120	48.88	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	97.76%		
47) 2-Hexanone-d5	8.091	63	116521	97.29	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	97.29%		
57) 1,1,2,2-Tetrachloroeth...	10.220	84	194844	48.71	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	97.42%		
64) 1,2-Dichlorobenzene-d4	11.628	152	164801	54.55	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	109.10%		
Target Compounds					Qvalue	
13) Acetone	2.182	43	9836	6.37	ug/L	83
25) Chloroform	4.381	83	14896	2.84	ug/L	96
48) 2-Hexanone	8.146	43	16149	4.82	ug/L #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031621\
Data File : VV020631.D
Acq On : 16 Mar 2021 15:26
Operator : SY/MD
Sample : M1675-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BC9

Quant Time: Mar 17 01:51:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM031321WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 17 01:47:28 2021
Response via : Initial Calibration

