

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV031623\  
 Data File : VV030222.D  
 Acq On : 16 Mar 2023 14:57  
 Operator : SY/MD  
 Sample : 01947-07  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 8 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_V  
**ClientSampleId :**  
 CB941

Quant Time: Mar 17 05:48:35 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR022323WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Fri Mar 17 05:46:57 2023  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	152101	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	145832	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	76206	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	37228	3.783	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	75.600%	
7) Chloroethane-d5	1.536	69	33424	3.618	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	72.400%	
11) 1,1-Dichloroethene-d2	2.063	63	58150	2.874	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	57.400%#	
20) 2-Butanone-d5	3.812	46	75756	41.750	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	83.500%	
24) Chloroform-d	4.256	84	73876	3.675	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	73.400%	
26) 1,2-Dichloroethane-d4	4.950	65	36380	4.433	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.600%	
32) Benzene-d6	4.966	84	136493	3.478	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	69.600%#	
36) 1,2-Dichloropropane-d6	5.995	67	42086	3.565	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	71.400%	
41) Toluene-d8	7.249	98	120219	3.232	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	64.600%#	
43) trans-1,3-Dichloroprop...	7.561	79	15278	3.783	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	75.600%	
46) 2-Hexanone-d5	8.034	63	57768	34.572	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	69.140%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	29321	3.491	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	69.800%	
66) 1,2-Dichlorobenzene-d4	11.570	152	44950	3.486	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	69.800%#	
Target Compounds						
13) Acetone	2.146	43	4265	3.180	ug/L	89
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV031623\  
 Data File : VV030222.D  
 Acq On : 16 Mar 2023 14:57  
 Operator : SY/MD  
 Sample : 01947-07  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 CB941

Quant Time: Mar 17 05:48:35 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR022323WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Fri Mar 17 05:46:57 2023  
 Response via : Initial Calibration

