

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031720\
 Data File : VV014966.D
 Acq On : 17 Mar 2020 12:04
 Operator : SY/MD
 Sample : VV0317WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK86

Quant Time: Mar 18 05:30:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031620WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 18 05:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	317403	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	314483	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	149182	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	87105	41.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.38%
7) Chloroethane-d5	1.58	69	72956	46.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.90%
11) 1,1-Dichloroethene-d2	2.12	63	116462	35.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.80%
21) 2-Butanone-d5	3.93	46	118994	92.29	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.29%
24) Chloroform-d	4.38	84	214483	48.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.36%
26) 1,2-Dichloroethane-d4	5.06	65	141450	50.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.36%
32) Benzene-d6	5.08	84	429686	49.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.20%
36) 1,2-Dichloropropane-d6	6.10	67	129445	50.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.36%
41) Toluene-d8	7.34	98	417974	48.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.56%
43) trans-1,3-Dichloropropene-	7.64	79	63519	46.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.62%
47) 2-Hexanone-d5	8.12	63	113550	97.51	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.51%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	185968	49.24	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.48%
64) 1,2-Dichlorobenzene-d4	11.65	152	174661	53.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.90%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031720\
Data File : VV014966.D
Acq On : 17 Mar 2020 12:04
Operator : SY/MD
Sample : VV0317WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK86

Quant Time: Mar 18 05:30:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031620WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 18 05:21:28 2020
Response via : Initial Calibration

