

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031720\
 Data File : VV014969.D
 Acq On : 17 Mar 2020 13:22
 Operator : SY/MD
 Sample : L1890-17DL 100X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DH4DL

Quant Time: Mar 18 05:44:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031620WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 18 05:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	341956	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	330251	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	164434	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	83739	36.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	73.52%
7) Chloroethane-d5	1.58	69	70417	41.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.24%
11) 1,1-Dichloroethene-d2	2.12	63	109726	31.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.78%
21) 2-Butanone-d5	3.93	46	106886	76.95	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	76.95%
24) Chloroform-d	4.38	84	208499	43.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.86%
26) 1,2-Dichloroethane-d4	5.06	65	141008	46.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.78%
32) Benzene-d6	5.08	84	423283	46.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.04%
36) 1,2-Dichloropropane-d6	6.10	67	128693	47.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.00%
41) Toluene-d8	7.34	98	411665	45.75	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.50%
43) trans-1,3-Dichloropropene-	7.64	79	62582	43.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.90%
47) 2-Hexanone-d5	8.12	63	110763	90.57	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.57%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	179566	45.27	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.54%
64) 1,2-Dichlorobenzene-d4	11.65	152	176553	49.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.02%

Target Compounds

					Ovalue
33) Benzene	5.13	78	1311761	145.971	ug/L 100
51) Chlorobenzene	8.90	112	1026019	147.345	ug/L 99
62) 1,3-Dichlorobenzene	11.20	146	5709	1.054	ug/L 92
63) 1,4-Dichlorobenzene	11.29	146	122442	22.432	ug/L 99
65) 1,2-Dichlorobenzene	11.67	146	64719	11.957	ug/L 97
68) 1,2,4-trichlorobenzene	13.28	180	5061	1.337	ug/L 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031720\
Data File : VV014969.D
Acq On : 17 Mar 2020 13:22
Operator : SY/MD
Sample : L1890-17DL 100X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DH4DL

Quant Time: Mar 18 05:44:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031620WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 18 05:21:28 2020
Response via : Initial Calibration

