

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031720\
 Data File : VV014970.D
 Acq On : 17 Mar 2020 13:46
 Operator : SY/MD
 Sample : L1890-18DL 200X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DH5DL

Quant Time: Mar 18 05:49:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031620WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 18 05:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	338829	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	327338	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	167280	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	81635	36.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.34%
7) Chloroethane-d5	1.58	69	68992	41.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.30%
11) 1,1-Dichloroethene-d2	2.12	63	109445	31.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.20%
21) 2-Butanone-d5	3.93	46	115426	83.86	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.86%
24) Chloroform-d	4.38	84	207112	44.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.08%
26) 1,2-Dichloroethane-d4	5.06	65	139118	46.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.38%
32) Benzene-d6	5.08	84	421335	46.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.44%
36) 1,2-Dichloropropane-d6	6.10	67	125483	46.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.46%
41) Toluene-d8	7.34	98	410667	46.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.10%
43) trans-1,3-Dichloropropene-	7.64	79	61717	43.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.46%
47) 2-Hexanone-d5	8.12	63	118126	97.45	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.45%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	184916	47.04	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.08%
64) 1,2-Dichlorobenzene-d4	11.65	152	178586	48.74	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.48%

Target Compounds

					Ovalue
33) Benzene	5.13	78	1664486	186.870	ug/L 100
51) Chlorobenzene	8.90	112	1660281	240.552	ug/L 99
62) 1,3-Dichlorobenzene	11.20	146	14421	2.616	ug/L 98
63) 1,4-Dichlorobenzene	11.29	146	383856	69.127	ug/L 99
65) 1,2-Dichlorobenzene	11.67	146	246719	44.805	ug/L 99
68) 1,2,4-trichlorobenzene	13.28	180	13604	3.531	ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031720\
Data File : VV014970.D
Acq On : 17 Mar 2020 13:46
Operator : SY/MD
Sample : L1890-18DL 200X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DH5DL

Quant Time: Mar 18 05:49:08 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031620WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 18 05:21:28 2020
Response via : Initial Calibration

