

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031720\
 Data File : VV014982.D
 Acq On : 17 Mar 2020 18:37
 Operator : SY/MD
 Sample : L1940-19
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DK2

Quant Time: Mar 18 07:03:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031620WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 18 05:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	432181	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	418443	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	203665	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	84202	29.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	58.50%#
7) Chloroethane-d5	1.58	69	69475	32.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	64.98%#
11) 1,1-Dichloroethene-d2	2.12	63	110958	25.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	50.24%#
21) 2-Butanone-d5	3.93	46	112548	64.11	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	64.11%
24) Chloroform-d	4.38	84	202500	33.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	67.52%#
26) 1,2-Dichloroethane-d4	5.06	65	138742	36.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	73.00%
32) Benzene-d6	5.08	84	415677	36.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	72.12%
36) 1,2-Dichloropropane-d6	6.10	67	124063	36.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	72.28%
41) Toluene-d8	7.34	98	400310	35.11	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	70.22%#
43) trans-1,3-Dichloropropene-	7.64	79	54333	29.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.54%#
47) 2-Hexanone-d5	8.12	63	115978	74.85	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.85%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	177748	35.37	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	70.74%
64) 1,2-Dichlorobenzene-d4	11.65	152	171187	38.37	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	76.74%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031720\
Data File : VV014982.D
Acq On : 17 Mar 2020 18:37
Operator : SY/MD
Sample : L1940-19
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DK2

Quant Time: Mar 18 07:03:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031620WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 18 05:21:28 2020
Response via : Initial Calibration

