

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031720\
 Data File : VV014983.D
 Acq On : 17 Mar 2020 19:01
 Operator : SY/MD
 Sample : L1941-01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DA1

Quant Time: Mar 18 07:05:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031620WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 18 05:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	321784	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	307689	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	150163	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	82322	38.40	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.80%
7) Chloroethane-d5	1.58	69	66149	41.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.10%
11) 1,1-Dichloroethene-d2	2.12	63	106805	32.47	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.94%
21) 2-Butanone-d5	3.93	46	107998	82.62	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	82.62%
24) Chloroform-d	4.38	84	199515	44.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.34%
26) 1,2-Dichloroethane-d4	5.06	65	133750	47.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.52%
32) Benzene-d6	5.08	84	409370	48.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.58%
36) 1,2-Dichloropropane-d6	6.10	67	122892	48.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.38%
41) Toluene-d8	7.34	98	390411	46.57	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.14%
43) trans-1,3-Dichloropropene-	7.64	79	52579	39.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.36%
47) 2-Hexanone-d5	8.12	63	110860	97.30	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.30%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	176284	47.70	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.40%
64) 1,2-Dichlorobenzene-d4	11.65	152	164291	49.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.88%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031720\
Data File : VV014983.D
Acq On : 17 Mar 2020 19:01
Operator : SY/MD
Sample : L1941-01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DA1

Quant Time: Mar 18 07:05:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031620WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 18 05:21:28 2020
Response via : Initial Calibration

