

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV031725\
Data File : VV038719.D
Acq On : 17 Mar 2025 16:44
Operator : SY/MD
Sample : VV0317WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Mar 18 08:16:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM031724WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Mar 17 15:08:07 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	369392	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	378090	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	159508	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	173600	46.68	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	93.36%	
7) Chloroethane-d5	1.529	69	146718	47.06	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	94.12%	
11) 1,1-Dichloroethene-d2	2.053	65	74642	44.76	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	89.52%	
21) 2-Butanone-d5	3.789	46	109104	81.44	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	81.44%	
24) Chloroform-d	4.239	84	308016	45.82	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.64%	
26) 1,2-Dichloroethane-d4	4.934	65	191887	49.06	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.12%	
32) Benzene-d6	4.950	84	604038	50.45	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.90%	
36) 1,2-Dichloropropane-d6	5.979	67	201546	52.14	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	104.28%	
41) Toluene-d8	7.236	98	496823	47.77	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	95.54%	
43) trans-1,3-Dichloroprop...	7.548	79	76134	47.79	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	95.58%	
47) 2-Hexanone-d5	8.021	63	72431	90.57	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	90.57%	
56) 1,1,2,2-Tetrachloroeth...	10.143	84	223584	46.64	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	93.28%	
66) 1,2-Dichlorobenzene-d4	11.551	152	171561	56.35	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	112.70%	

Target Compounds

ABTKS00
Client Sample :
MSVOA_V
Instrument :

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV031725\
Data File : VV038719.D
Acq On : 17 Mar 2025 16:44
Operator : SY/MD
Sample : VV0317WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Mar 18 08:16:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM031724WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Mar 17 15:08:07 2025
Response via : Initial Calibration

