

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031819\
 Data File : VV009771.D
 Acq On : 18 Mar 2019 20:37
 Operator : SY/MD
 Sample : I.BLK
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK41

Quant Time: Mar 19 05:45:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Mar 18 12:32:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	144205	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	136043	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	52202	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	39485	40.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.22%
7) Chloroethane-d5	1.56	69	33223	46.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.64%
11) 1,1-Dichloroethene-d2	2.12	63	66387	32.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.12%
21) 2-Butanone-d5	3.93	46	53316	92.06	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.06%
24) Chloroform-d	4.40	84	77890	43.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.22%
26) 1,2-Dichloroethane-d4	5.08	65	50794	46.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.70%
32) Benzene-d6	5.10	84	166041	46.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.42%
36) 1,2-Dichloropropane-d6	6.12	67	50641	46.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.58%
41) Toluene-d8	7.36	98	155821	45.55	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.10%
43) trans-1,3-Dichloropropene-	7.66	79	22334	43.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.28%
47) 2-Hexanone-d5	8.13	63	42944	97.49	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.49%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	72409	44.10	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	50752	48.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.10%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031819\
Data File : VV009771.D
Acq On : 18 Mar 2019 20:37
Operator : SY/MD
Sample : I.BLK
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK41

Quant Time: Mar 19 05:45:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031719WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Mar 18 12:32:58 2019
Response via : Initial Calibration

