

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031820\
 Data File : VV015028.D
 Acq On : 18 Mar 2020 21:12
 Operator : SY/MD
 Sample : L1941-05DL 40X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DB5DL

Quant Time: Mar 19 05:14:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 19 04:08:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	334666	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	327952	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	164733	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	84517	54.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	108.10%
7) Chloroethane-d5	1.58	69	66899	54.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.12%
11) 1,1-Dichloroethene-d2	2.12	63	104791	39.21	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.42%
21) 2-Butanone-d5	3.93	46	116003	109.43	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.43%
24) Chloroform-d	4.38	84	200698	50.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.26%
26) 1,2-Dichloroethane-d4	5.06	65	132148	52.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.88%
32) Benzene-d6	5.08	84	415743	53.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.84%
36) 1,2-Dichloropropane-d6	6.10	67	120722	53.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.04%
41) Toluene-d8	7.34	98	405571	51.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.54%
43) trans-1,3-Dichloropropene-	7.64	79	59687	49.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.08%
47) 2-Hexanone-d5	8.12	63	111140	105.92	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.92%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	171535	50.06	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.12%
64) 1,2-Dichlorobenzene-d4	11.65	152	175073	52.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.88%

Target Compounds

					Ovalue
33) Benzene	5.13	78	6392	0.796 ug/L	100
51) Chlorobenzene	8.90	112	24047	3.769 ug/L	98
62) 1,3-Dichlorobenzene	11.20	146	22166	4.470 ug/L	96
63) 1,4-Dichlorobenzene	11.29	146	330427	64.019 ug/L	98
65) 1,2-Dichlorobenzene	11.66	146	94030	17.854 ug/L	99
68) 1,2,4-trichlorobenzene	13.28	180	23668	6.551 ug/L	97
70) 1,2,3-Trichlorobenzene	13.77	180	4775	1.345 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031820\
Data File : VV015028.D
Acq On : 18 Mar 2020 21:12
Operator : SY/MD
Sample : L1941-05DL 40X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DB5DL

Quant Time: Mar 19 05:14:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 19 04:08:09 2020
Response via : Initial Calibration

