

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031820\
 Data File : VV015030.D
 Acq On : 18 Mar 2020 22:01
 Operator : SY/MD
 Sample : L1940-19
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0DK2

Quant Time: Mar 19 05:23:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 19 04:08:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	336942	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	325360	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	159966	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	83462	53.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	106.04%
7) Chloroethane-d5	1.58	69	65416	52.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.98%
11) 1,1-Dichloroethene-d2	2.12	63	106180	39.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.92%
21) 2-Butanone-d5	3.93	46	105325	98.68	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.68%
24) Chloroform-d	4.38	84	197291	48.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.90%
26) 1,2-Dichloroethane-d4	5.06	65	130890	51.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.18%
32) Benzene-d6	5.08	84	408154	52.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.72%
36) 1,2-Dichloropropane-d6	6.10	67	116832	51.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.44%
41) Toluene-d8	7.34	98	400999	51.59	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.18%
43) trans-1,3-Dichloropropene-	7.64	79	58474	48.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.84%
47) 2-Hexanone-d5	8.12	63	107681	103.45	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.45%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	172842	50.85	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.70%
64) 1,2-Dichlorobenzene-d4	11.64	152	169284	52.71	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.42%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031820\
Data File : VV015030.D
Acq On : 18 Mar 2020 22:01
Operator : SY/MD
Sample : L1940-19
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0DK2

Quant Time: Mar 19 05:23:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 19 04:08:09 2020
Response via : Initial Calibration

