

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031820\
 Data File : VV015042.D
 Acq On : 19 Mar 2020 02:52
 Operator : SY/MD
 Sample : VIBLK72
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK72

Quant Time: Mar 19 07:06:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 19 04:08:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	310301	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	292472	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	136878	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	100324	69.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	138.40%#
7) Chloroethane-d5	1.58	69	68164	59.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	119.92%
11) 1,1-Dichloroethene-d2	2.12	63	97498	39.35	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.70%
21) 2-Butanone-d5	3.94	46	98227	99.93	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.93%
24) Chloroform-d	4.38	84	196771	53.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.02%
26) 1,2-Dichloroethane-d4	5.06	65	132055	56.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	113.04%
32) Benzene-d6	5.08	84	397338	57.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.50%
36) 1,2-Dichloropropane-d6	6.09	67	119175	58.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	117.38%
41) Toluene-d8	7.34	98	369155	52.84	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.68%
43) trans-1,3-Dichloropropene-	7.64	79	60227	55.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	110.96%
47) 2-Hexanone-d5	8.12	63	112420	120.14	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	120.14%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	173483	56.77	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	113.54%
64) 1,2-Dichlorobenzene-d4	11.64	152	158360	57.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	115.26%

Target Compounds

					Ovalue
33) Benzene	5.13	78	18784	2.624 ug/L	100
51) Chlorobenzene	8.90	112	102495	18.011 ug/L	97
62) 1,3-Dichlorobenzene	11.20	146	27520	6.680 ug/L	96
63) 1,4-Dichlorobenzene	11.29	146	179590	41.876 ug/L	98
65) 1,2-Dichlorobenzene	11.66	146	166260	37.992 ug/L	99
68) 1,2,4-trichlorobenzene	13.28	180	85671	28.538 ug/L	99
70) 1,2,3-Trichlorobenzene	13.76	180	33540	11.372 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031820\
Data File : VV015042.D
Acq On : 19 Mar 2020 02:52
Operator : SY/MD
Sample : VIBLK72
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK72

Quant Time: Mar 19 07:06:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 19 04:08:09 2020
Response via : Initial Calibration

