

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031820\
 Data File : VV015022.D
 Acq On : 18 Mar 2020 18:40
 Operator : SY/MD
 Sample : VV0318WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK88

Quant Time: Mar 19 04:34:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 19 04:08:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	377347	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	363707	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	180569	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	71872	40.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.52%
7) Chloroethane-d5	1.58	69	58142	42.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.10%
11) 1,1-Dichloroethene-d2	2.12	63	91722	30.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.88%
21) 2-Butanone-d5	3.93	46	105782	88.50	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.50%
24) Chloroform-d	4.38	84	182011	40.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.64%
26) 1,2-Dichloroethane-d4	5.06	65	119560	42.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.16%
32) Benzene-d6	5.08	84	373672	43.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.58%
36) 1,2-Dichloropropane-d6	6.10	67	111044	43.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.94%
41) Toluene-d8	7.34	98	365776	42.10	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.20%
43) trans-1,3-Dichloropropene-	7.64	79	55271	40.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.88%
47) 2-Hexanone-d5	8.12	63	109054	93.72	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.72%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	162675	42.81	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.62%
64) 1,2-Dichlorobenzene-d4	11.65	152	160132	44.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.36%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031820\
Data File : VV015022.D
Acq On : 18 Mar 2020 18:40
Operator : SY/MD
Sample : VV0318WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK88

Quant Time: Mar 19 04:34:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031820WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 19 04:08:09 2020
Response via : Initial Calibration

