

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV031919\
 Data File : VV009786.D
 Acq On : 19 Mar 2019 18:31
 Operator : SY/MD
 Sample : K1986-18
 Misc : 5.90u/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BF126

Quant Time: Mar 20 06:35:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 20 05:16:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	136942	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	137318	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	58013	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	38631	41.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.68%
7) Chloroethane-d5	1.55	69	20516	30.12	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	60.24%#
11) 1,1-Dichloroethene-d2	2.10	63	66797	33.97	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.94%
21) 2-Butanone-d5	3.95	46	49961	90.84	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.84%
24) Chloroform-d	4.40	84	78395	46.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.44%
26) 1,2-Dichloroethane-d4	5.08	65	50368	48.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.84%
32) Benzene-d6	5.09	84	169162	47.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.28%
36) 1,2-Dichloropropane-d6	6.12	67	51651	47.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.56%
41) Toluene-d8	7.36	98	157946	45.75	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.50%
43) trans-1,3-Dichloropropene-	7.67	79	23323	45.15	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.30%
47) 2-Hexanone-d5	8.14	63	42234	94.99	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.99%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	73465	44.33	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.66%
64) 1,2-Dichlorobenzene-d4	11.68	152	57144	48.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.36%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV031919\
Data File : VV009786.D
Acq On : 19 Mar 2019 18:31
Operator : SY/MD
Sample : K1986-18
Misc : 5.90q/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF126

Quant Time: Mar 20 06:35:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM031719WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 20 05:16:13 2019
Response via : Initial Calibration

