

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031924\
 Data File : VV034712.D
 Acq On : 18 Mar 2024 21:55
 Operator : SY/MD
 Sample : P1766-15
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Mar 19 02:57:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 19 02:53:17 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	328214	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	311785	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	143869	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	156710	55.077	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 110.160%		
7) Chloroethane-d5	1.529	69	121487	54.806	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 109.620%		
11) 1,1-Dichloroethene-d2	2.060	65	67331	52.326	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 104.660%		
21) 2-Butanone-d5	3.789	46	189911	99.809	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 99.810%		
24) Chloroform-d	4.246	84	259100	51.503	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 103.000%		
26) 1,2-Dichloroethane-d4	4.940	65	173621	52.813	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 105.620%		
32) Benzene-d6	4.960	84	510382	54.240	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 108.480%		
36) 1,2-Dichloropropane-d6	5.985	67	156904	54.638	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 109.280%		
41) Toluene-d8	7.243	98	447825	52.590	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 105.180%		
43) trans-1,3-Dichloroprop...	7.551	79	70425	48.836	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 97.680%		
47) 2-Hexanone-d5	8.024	63	120413	83.971	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 83.970%		
56) 1,1,2,2-Tetrachloroeth...	10.152	84	202935	53.700	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 107.400%		
66) 1,2-Dichlorobenzene-d4	11.561	152	165091	56.140	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 112.280%		
Target Compounds						
					Qvalue	
10) 1,1,2-Trichloro-1,2,2-...	2.063	101	1963	0.741	ug/L #	18
12) 1,1-Dichloroethene	2.069	96	13148	5.350	ug/L #	1
13) Acetone	2.121	43	10131	4.429	ug/L	98
20) cis-1,2-Dichloroethene	3.818	96	16549	6.412	ug/L	97
30) 1,1,1-Trichloroethane	4.513	97	8809	1.970	ug/L	96
34) Trichloroethene	5.831	95	152405	55.401	ug/L	97
61) 1,2,3-Trichloropropane	11.185	75	17226	6.272	ug/L #	60


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031924\
Data File : VV034712.D
Acq On : 18 Mar 2024 21:55
Operator : SY/MD
Sample : P1766-15
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Mar 19 02:57:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 19 02:53:17 2024
Response via : Initial Calibration





#10

1,1,2-Trichloro-1,2,2-trifluoroethane

Concen: 0.741 ug/L

RT: 2.063 min Scan# 319

Delta R.T. -0.003 min

Lab File: VV034712.D





Acq: 18 Mar 2024 21:55

Instrument :

MSVOA_V

ClientSampleId :

Tgt Ion:	101	Resp:	1963
Ion	Ratio	Lower	Upper
101	100		
85	0.0	37.4	56.0#
151	0.0	61.4	92.2#



#12

1,1-Dichloroethene

Concen: 5.350 ug/L




RT: 2.069 min Scan# 320


Delta R.T. 0.000 min

Lab File: VV034712.D

Acq: 18 Mar 2024 21:55

Tgt Ion:	96	Resp:	13148
Ion	Ratio	Lower	Upper
96	100		
61	274.2	128.9	239.3#
63	1000.2	98.0	182.0#





#13

Acetone

Concen: 4.429 ug/L

RT: 2.121 min Scan# 335

Delta R.T. 0.003 min

Lab File: VV034712.D

Acq: 18 Mar 2024 21:55

Instrument :

MSVOA_V





ClientSampleId :

Tgt Ion: 43 Resp: 10131

Ion Ratio Lower Upper

43 100

58 29.5 0.0 57.2



#20

cis-1,2-Dichloroethene

Concen: 6.412 ug/L

RT: 3.818 min Scan# 864

Delta R.T. 0.006 min

Lab File: VV034712.D

Acq: 18 Mar 2024 21:55




Tgt Ion: 96 Resp: 16549


Ion Ratio Lower Upper

96 100

61 133.4 96.0 178.4

98 68.2 46.0 85.4





#30

1,1,1-Trichloroethane

Concen: 1.970 ug/L

RT: 4.513 min Scan# 1

Delta R.T. 0.003 min

Lab File: VV034712.D

Acq: 18 Mar 2024 21:55

Instrument :

MSVOA_V

ClientSampleId :





Tgt Ion: 97 Resp: 8809

Ion Ratio Lower Upper

97 100

99 66.1 51.3 76.9

61 49.6 36.0 54.0



#34

Trichloroethene

Concen: 55.401 ug/L

RT: 5.831 min Scan# 1490

Delta R.T. 0.000 min

Lab File: VV034712.D

Acq: 18 Mar 2024 21:55

Tgt Ion: 95 Resp: 152405




Ion Ratio Lower Upper


95 100

97 66.2 44.7 83.1

132 95.3 65.3 121.3


130 102.6 69.1 128.3





#61
 1,2,3-Trichloropropane
 Concen: 6.272 ug/L
 RT: 11.185 min Scan# 3155
 Delta R.T. 0.978 min
 Lab File: VV034712.D
 Acq: 18 Mar 2024 21:55

Instrument :
 MSVOA_V
 ClientSampleId :



Tgt Ion: 75 Resp: 17226
 Ion Ratio Lower Upper
 75 100
 77 36.9 25.9 38.9
 110 1.1 33.1 49.7#

