

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031924\
 Data File : VV034722.D
 Acq On : 19 Mar 2024 10:02
 Operator : SY/MD
 Sample : VV0319WBL02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK249

Quant Time: Mar 20 01:20:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 20 01:19:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	359865	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	339072	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	159626	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	164819	52.832	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 105.660%		
7) Chloroethane-d5	1.536	69	130532	53.707	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 107.420%		
11) 1,1-Dichloroethene-d2	2.063	65	73220	51.897	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 103.800%		
21) 2-Butanone-d5	3.793	46	199306	95.534	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 95.530%		
24) Chloroform-d	4.249	84	278441	50.480	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 100.960%		
26) 1,2-Dichloroethane-d4	4.941	65	187252	51.950	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 103.900%		
32) Benzene-d6	4.960	84	554974	54.232	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 108.460%		
36) 1,2-Dichloropropane-d6	5.986	67	167360	53.589	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 107.180%		
41) Toluene-d8	7.240	98	492849	53.220	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 106.440%		
43) trans-1,3-Dichloroprop...	7.551	79	80562	51.370	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 102.740%		
47) 2-Hexanone-d5	8.024	63	124655	79.933	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 79.930%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	213638	51.982	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 103.960%		
66) 1,2-Dichlorobenzene-d4	11.558	152	177933	54.534	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 109.060%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031924\
Data File : VV034722.D
Acq On : 19 Mar 2024 10:02
Operator : SY/MD
Sample : VV0319WBL02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK249

Quant Time: Mar 20 01:20:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 20 01:19:21 2024
Response via : Initial Calibration

