

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031924\
 Data File : VW034727.D
 Acq On : 19 Mar 2024 12:00
 Operator : SY/MD
 Sample : P1752-01DL 5X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCL01DL

Quant Time: Mar 20 01:22:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 20 01:19:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	346315	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	325313	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	147422	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	147720	49.204	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.400%
7) Chloroethane-d5	1.532	69	117385	50.188	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.380%
11) 1,1-Dichloroethene-d2	2.060	65	64983	47.861	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.720%
21) 2-Butanone-d5	3.789	46	186946	93.115	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.120%
24) Chloroform-d	4.246	84	259065	48.805	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.600%
26) 1,2-Dichloroethane-d4	4.940	65	173475	50.010	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.020%
32) Benzene-d6	4.957	84	506319	51.570	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.140%
36) 1,2-Dichloropropane-d6	5.985	67	155033	51.742	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.480%
41) Toluene-d8	7.239	98	445789	50.174	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.340%
43) trans-1,3-Dichloroprop...	7.551	79	74362	49.422	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.840%
47) 2-Hexanone-d5	8.024	63	114350	76.427	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.430%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	198632	50.375	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.760%
66) 1,2-Dichlorobenzene-d4	11.561	152	162757	54.012	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.020%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.069	96	118628	45.746	ug/L	87
13) Acetone	2.124	43	59617	24.701	ug/L	98
19) 1,1-Dichloroethane	3.117	63	18139	3.689	ug/L	96
27) 1,2-Dichloroethane	5.059	62	8620	2.082	ug/L	99
30) 1,1,1-Trichloroethane	4.513	97	33091	7.094	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031924\
Data File : VV034727.D
Acq On : 19 Mar 2024 12:00
Operator : SY/MD
Sample : P1752-01DL 5X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCL01DL

Quant Time: Mar 20 01:22:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 20 01:19:21 2024
Response via : Initial Calibration

