

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW031924\  
 Data File : VW034743.D  
 Acq On : 19 Mar 2024 18:21  
 Operator : SY/MD  
 Sample : P1755-03  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 YCKT5

Quant Time: Mar 20 01:26:25 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM022924WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Wed Mar 20 01:19:21 2024  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	283117	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	272209	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	120825	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	106171	43.259	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery =	86.520%		
7) Chloroethane-d5	1.523	69	91432	47.818	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery =	95.640%		
11) 1,1-Dichloroethene-d2	2.057	65	47500	42.794	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery =	85.580%		
21) 2-Butanone-d5	3.786	46	171584	104.541	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery =	104.540%		
24) Chloroform-d	4.246	84	211925	48.836	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	97.680%		
26) 1,2-Dichloroethane-d4	4.941	65	146403	51.627	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	103.260%		
32) Benzene-d6	4.957	84	408437	49.716	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	99.440%		
36) 1,2-Dichloropropane-d6	5.986	67	129746	51.750	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery =	103.500%		
41) Toluene-d8	7.243	98	356363	47.934	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	95.860%		
43) trans-1,3-Dichloroprop...	7.551	79	61281	48.674	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery =	97.340%		
47) 2-Hexanone-d5	8.024	63	102234	81.659	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery =	81.660%		
56) 1,1,2,2-Tetrachloroeth...	10.153	84	173491	52.583	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery =	105.160%		
66) 1,2-Dichlorobenzene-d4	11.561	152	130996	53.041	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	106.080%		
Target Compounds						
13) Acetone	2.114	43	56022	28.393	ug/L	98
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV031924\  
 Data File : VV034743.D  
 Acq On : 19 Mar 2024 18:21  
 Operator : SY/MD  
 Sample : P1755-03  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 25 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 YCKT5

Quant Time: Mar 20 01:26:25 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM022924WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Wed Mar 20 01:19:21 2024  
 Response via : Initial Calibration

