

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031924\
 Data File : VW034708.D
 Acq On : 18 Mar 2024 20:19
 Operator : SY/MD
 Sample : P1766-11
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COGM9

Quant Time: Mar 19 02:57:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 19 02:53:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	329716	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	310851	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	141296	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	153880	53.836	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 107.680%		
7) Chloroethane-d5	1.529	69	120825	54.259	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 108.520%		
11) 1,1-Dichloroethene-d2	2.056	65	65488	50.661	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 101.320%		
21) 2-Butanone-d5	3.786	46	199812	104.534	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 104.530%		
24) Chloroform-d	4.246	84	253650	50.190	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 100.380%		
26) 1,2-Dichloroethane-d4	4.941	65	172050	52.097	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 104.200%		
32) Benzene-d6	4.960	84	493814	52.637	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 105.280%		
36) 1,2-Dichloropropane-d6	5.986	67	152247	53.176	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 106.360%		
41) Toluene-d8	7.243	98	443083	52.189	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 104.380%		
43) trans-1,3-Dichloroprop...	7.555	79	69089	48.054	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 96.100%		
47) 2-Hexanone-d5	8.024	63	121057	84.673	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 84.670%		
56) 1,1,2,2-Tetrachloroeth...	10.153	84	201787	53.556	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 107.120%		
66) 1,2-Dichlorobenzene-d4	11.561	152	160292	55.500	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 111.000%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031924\
Data File : VV034708.D
Acq On : 18 Mar 2024 20:19
Operator : SY/MD
Sample : P1766-11
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0GM9

Quant Time: Mar 19 02:57:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 19 02:53:17 2024
Response via : Initial Calibration

