

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031924\
 Data File : VW034713.D
 Acq On : 18 Mar 2024 22:19
 Operator : SY/MD
 Sample : P1766-16
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COGN7

Quant Time: Mar 19 02:58:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 19 02:53:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	332553	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	313612	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	143203	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	152166	52.783	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.560%
7) Chloroethane-d5	1.529	69	120547	53.673	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.340%
11) 1,1-Dichloroethene-d2	2.060	65	66355	50.894	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.780%
21) 2-Butanone-d5	3.786	46	200874	104.193	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.190%
24) Chloroform-d	4.246	84	253568	49.746	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.500%
26) 1,2-Dichloroethane-d4	4.941	65	167723	50.353	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.700%
32) Benzene-d6	4.960	84	498098	52.626	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.260%
36) 1,2-Dichloropropane-d6	5.986	67	152907	52.936	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.880%
41) Toluene-d8	7.243	98	439161	51.272	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.540%
43) trans-1,3-Dichloroprop...	7.551	79	69605	47.986	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.980%
47) 2-Hexanone-d5	8.024	63	119851	83.092	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.090%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	201031	52.886	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.780%
66) 1,2-Dichlorobenzene-d4	11.561	152	160932	54.980	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031924\
Data File : VV034713.D
Acq On : 18 Mar 2024 22:19
Operator : SY/MD
Sample : P1766-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COGN7

Quant Time: Mar 19 02:58:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 19 02:53:17 2024
Response via : Initial Calibration

