

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031924\
 Data File : VW034730.D
 Acq On : 19 Mar 2024 13:12
 Operator : SY/MD
 Sample : P1751-15
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCL32

Quant Time: Mar 20 01:22:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 20 01:19:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	346989	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	326284	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	151753	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	144326	47.980	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.960%
7) Chloroethane-d5	1.532	69	117798	50.267	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.540%
11) 1,1-Dichloroethene-d2	2.059	65	64015	47.057	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.120%
21) 2-Butanone-d5	3.789	46	182023	90.487	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.490%
24) Chloroform-d	4.246	84	254405	47.834	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.660%
26) 1,2-Dichloroethane-d4	4.940	65	169085	48.650	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.300%
32) Benzene-d6	4.956	84	493017	50.066	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.140%
36) 1,2-Dichloropropane-d6	5.985	67	152139	50.625	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.240%
41) Toluene-d8	7.243	98	442386	49.643	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.280%
43) trans-1,3-Dichloroprop...	7.551	79	72444	48.004	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.000%
47) 2-Hexanone-d5	8.024	63	113501	75.633	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.630%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	196257	49.625	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.240%
66) 1,2-Dichlorobenzene-d4	11.557	152	161086	51.932	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.860%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.069	96	18147	6.984	ug/L #	1
13) Acetone	2.124	43	95059	39.309	ug/L	98
30) 1,1,1-Trichloroethane	4.513	97	13513	2.888	ug/L #	85


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031924\
Data File : VV034730.D
Acq On : 19 Mar 2024 13:12
Operator : SY/MD
Sample : P1751-15
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCL32

Quant Time: Mar 20 01:22:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 20 01:19:21 2024
Response via : Initial Calibration





#12

1,1-Dichloroethene

Concen: 6.984 ug/L

RT: 2.069 min Scan# 31

Delta R.T. -0.000 min

Lab File: VV034730.D

Acq: 19 Mar 2024 13:12

Instrument :

MSVOA_V

ClientSampleId :

YCL32





Tgt Ion: 96 Resp: 18147

Ion Ratio Lower Upper

96 100

61 261.7 128.9 239.3#

63 778.5 98.0 182.0#



#13

Acetone

Concen: 39.309 ug/L

RT: 2.124 min Scan# 337

Delta R.T. 0.003 min

Lab File: VV034730.D




Acq: 19 Mar 2024 13:12


Tgt Ion: 43 Resp: 95059

Ion Ratio Lower Upper

43 100

58 29.7 0.0 57.2





#30
 1,1,1-Trichloroethane
 Concen: 2.888 ug/L
 RT: 4.513 min Scan# 10
 Delta R.T. 0.006 min
 Lab File: VV034730.D
 Acq: 19 Mar 2024 13:12

Instrument :
 MSVOA_V
 ClientSampleId :
 YCL32

Tgt Ion: 97	Resp:	13513
Ion	Ratio	Lower Upper
97	100	
99	61.2	51.3 76.9
61	24.4	36.0 54.0

