

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW031924\
 Data File : VW034739.D
 Acq On : 19 Mar 2024 16:45
 Operator : SY/MD
 Sample : P1756-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCL59

Quant Time: Mar 20 01:25:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 20 01:19:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	330875	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	308747	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	142824	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	134133	46.763	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.520%
7) Chloroethane-d5	1.529	69	111981	50.111	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.220%
11) 1,1-Dichloroethene-d2	2.060	65	60066	46.304	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.600%
21) 2-Butanone-d5	3.790	46	179478	93.567	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.570%
24) Chloroform-d	4.249	84	241012	47.523	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.040%
26) 1,2-Dichloroethane-d4	4.941	65	161913	48.855	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.720%
32) Benzene-d6	4.960	84	470274	50.469	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.940%
36) 1,2-Dichloropropane-d6	5.989	67	143846	50.584	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.160%
41) Toluene-d8	7.243	98	408087	48.395	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.780%
43) trans-1,3-Dichloroprop...	7.555	79	68943	48.279	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.560%
47) 2-Hexanone-d5	8.024	63	108110	76.133	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.130%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	184401	49.275	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.560%
66) 1,2-Dichlorobenzene-d4	11.561	152	153837	52.695	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV031924\
Data File : VV034739.D
Acq On : 19 Mar 2024 16:45
Operator : SY/MD
Sample : P1756-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCL59

Quant Time: Mar 20 01:25:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 20 01:19:21 2024
Response via : Initial Calibration

