

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032023\
 Data File : VW030249.D
 Acq On : 20 Mar 2023 11:29
 Operator : SY/MD
 Sample : 02005-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COM99

Quant Time: Mar 20 12:35:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 17 22:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	120528	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	121527	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	57665	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	29980	3.844	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.800%
7) Chloroethane-d5	1.536	69	30486	4.164	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.200%
11) 1,1-Dichloroethene-d2	2.063	63	53350	3.327	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.600%
20) 2-Butanone-d5	3.838	46	76086	52.916	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.840%
24) Chloroform-d	4.262	84	71536	4.490	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
26) 1,2-Dichloroethane-d4	4.954	65	37013	5.691	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.800%
32) Benzene-d6	4.973	84	125655	3.843	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.800%
36) 1,2-Dichloropropane-d6	5.998	67	39651	4.031	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.600%
41) Toluene-d8	7.252	98	109909	3.546	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.000%
43) trans-1,3-Dichloroprop...	7.561	79	14639	4.349	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.037	63	61600	44.239	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.480%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	30398	4.343	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.800%
66) 1,2-Dichlorobenzene-d4	11.571	152	44740	4.586	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.800%
Target Compounds						
3) Chloromethane	1.217	50	4236	0.274	ug/L	96
16) Methylene chloride	2.455	84	2365	0.239	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032023\
Data File : VV030249.D
Acq On : 20 Mar 2023 11:29
Operator : SY/MD
Sample : 02005-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COM99

Quant Time: Mar 20 12:35:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 17 22:49:08 2023
Response via : Initial Calibration

