

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032023\
 Data File : VW030264.D
 Acq On : 20 Mar 2023 17:56
 Operator : SY/MD
 Sample : 01947-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CB939

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 03/21/2023
 Supervised By :Mahesh Dadoda 03/21/2023

Quant Time: Mar 21 02:58:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 17 22:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	129622	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	127759	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	66702	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	25277	3.014	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	60.200%
7) Chloroethane-d5	1.539	69	25990	3.301	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	66.000%
11) 1,1-Dichloroethene-d2	2.066	63	45061	2.613	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	52.200%#
20) 2-Butanone-d5	3.818	46	66269	42.855	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.720%
24) Chloroform-d	4.262	84	64469	3.763	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.200%
26) 1,2-Dichloroethane-d4	4.957	65	33244	4.753	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.000%
32) Benzene-d6	4.973	84	111973	3.257	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	65.200%#
36) 1,2-Dichloropropane-d6	5.998	67	34871	3.372	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	67.400%
41) Toluene-d8	7.252	98	99513	3.054	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	61.000%#
43) trans-1,3-Dichloroprop...	7.564	79	12854	3.633	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.600%
46) 2-Hexanone-d5	8.034	63	54398	37.161	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.320%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	26936	3.661	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	73.200%
66) 1,2-Dichlorobenzene-d4	11.570	152	44577	3.950	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	79.000%#
Target Compounds					Qvalue	
13) Acetone	2.143	43	8318m	7.278	ug/L	
17) Methyl tert-butyl Ether	2.712	73	6884	0.405	ug/L #	81
33) Benzene	5.027	78	4337	0.124	ug/L	100
42) Toluene	7.330	91	3871	0.102	ug/L	94
51) Chlorobenzene	8.821	112	49593	1.946	ug/L	96
53) m,p-Xylene	9.085	106	2184	0.138	ug/L	92
54) o-Xylene	9.487	106	1693	0.112	ug/L	92
65) 1,4-Dichlorobenzene	11.220	146	2380	0.118	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032023\
Data File : VV030264.D
Acq On : 20 Mar 2023 17:56
Operator : SY/MD
Sample : 01947-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CB939

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 03/21/2023
Supervised By :Mahesh Dadoda 03/21/2023

Quant Time: Mar 21 02:58:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 17 22:49:08 2023
Response via : Initial Calibration

