

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032119\
 Data File : VV009809.D
 Acq On : 21 Mar 2019 21:16
 Operator : SY/MD
 Sample : K2049-10
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 P015-TW001-01

Quant Time: Mar 22 10:06:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR032119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 22 08:14:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	88582	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	91672	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	41312	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	57104	5.33	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.60%
7) Chloroethane-d5	1.59	69	52777	5.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.40%
11) 1,1-Dichloroethene-d2	2.13	63	108034	4.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.00%
20) 2-Butanone-d5	3.98	46	71930	58.06	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.12%
24) Chloroform-d	4.40	84	58974	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.60%
26) 1,2-Dichloroethane-d4	5.09	65	28962	5.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.60%
32) Benzene-d6	5.10	84	126073	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.80%
36) 1,2-Dichloropropane-d6	6.12	67	37380	5.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.40%
41) Toluene-d8	7.36	98	112965	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.00%
43) trans-1,3-Dichloropropene-	7.67	79	11588	4.37	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.40%
46) 2-Hexanone-d5	8.14	63	62352	55.33	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.66%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	27639	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	40631	5.41	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032119\
Data File : VV009809.D
Acq On : 21 Mar 2019 21:16
Operator : SY/MD
Sample : K2049-10
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
P015-TW001-01

Quant Time: Mar 22 10:06:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR032119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 22 08:14:34 2019
Response via : Initial Calibration

