

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032119\
 Data File : VV009804.D
 Acq On : 21 Mar 2019 19:07
 Operator : SY/MD
 Sample : K2049-05
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 P003-TW001-01

Quant Time: Mar 22 10:04:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR032119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 22 08:14:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	93276	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	96979	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	42215	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	54173	4.81	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.20%
7) Chloroethane-d5	1.59	69	52384	5.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.80%
11) 1,1-Dichloroethene-d2	2.13	63	104712	3.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.40%
20) 2-Butanone-d5	3.95	46	69839	53.53	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.06%
24) Chloroform-d	4.41	84	58735	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	5.09	65	30378	5.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.20%
32) Benzene-d6	5.10	84	127387	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.00%
36) 1,2-Dichloropropane-d6	6.12	67	38233	5.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.80%
41) Toluene-d8	7.36	98	117095	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.00%
43) trans-1,3-Dichloropropene-	7.67	79	12386	4.42	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.40%
46) 2-Hexanone-d5	8.14	63	62214	52.19	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.38%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	28219	4.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	40157	5.24	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032119\
Data File : VV009804.D
Acq On : 21 Mar 2019 19:07
Operator : SY/MD
Sample : K2049-05
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
P003-TW001-01

Quant Time: Mar 22 10:04:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR032119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 22 08:14:34 2019
Response via : Initial Calibration

