

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032124\
 Data File : VV034777.D
 Acq On : 21 Mar 2024 10:25
 Operator : SY/MD
 Sample : VV0321WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK251

Quant Time: Mar 22 02:00:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 22 01:59:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	345059	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	323191	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	151149	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	164739	55.073	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 110.140%		
7) Chloroethane-d5	1.532	69	127831	54.853	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 109.700%		
11) 1,1-Dichloroethene-d2	2.060	65	59507	43.988	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 87.980%		
21) 2-Butanone-d5	3.793	46	188247	94.105	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 94.100%		
24) Chloroform-d	4.246	84	268103	50.691	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 101.380%		
26) 1,2-Dichloroethane-d4	4.940	65	182990	52.946	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 105.900%		
32) Benzene-d6	4.957	84	535959	54.948	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 109.900%		
36) 1,2-Dichloropropane-d6	5.985	67	160829	54.029	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 108.060%		
41) Toluene-d8	7.239	98	471366	53.401	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 106.800%		
43) trans-1,3-Dichloroprop...	7.551	79	75045	50.203	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 100.400%		
47) 2-Hexanone-d5	8.024	63	111717	75.157	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 75.160%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	200477	51.177	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 102.360%		
66) 1,2-Dichlorobenzene-d4	11.558	152	169067	54.723	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 109.440%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032124\
Data File : VV034777.D
Acq On : 21 Mar 2024 10:25
Operator : SY/MD
Sample : VV0321WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK251

Quant Time: Mar 22 02:00:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 22 01:59:49 2024
Response via : Initial Calibration

